



UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/731,422	12/08/2003	Alexandra Naughton	279.B41US1	3939
21186 7590 07/18/2007 SCHWEGMAN, LUNDBERG, WOESSNER & KLUTH, P.A. P.O. BOX 2938 MINNEAPOLIS, MN 55402				
			EXAMINER MANUEL, GEORGE C	
			ART UNIT 3762	PAPER NUMBER
			MAIL DATE 07/18/2007	DELIVERY MODE PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

Office Action Summary

Application No.

10/731,422

Applicant(s)

NAUGHTON ET AL.

Examiner

George Manuel

Art Unit

3762

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 07 May 2007.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-17 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-17 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
 - ☐ Certified copies of the priority documents have been received in Application No. _____.
 - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
- * See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | Paper No(s)/Mail Date: _____ |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO/SB/08) | 5) <input type="checkbox"/> Notice of Informal Patent Application |
| Paper No(s)/Mail Date: _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless --

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

(e) the invention was described in a patent granted on an application for patent by another filed in the United States before the invention thereof by the applicant for patent, or on an international application by another who has fulfilled the requirements of paragraphs (1), (2), and (4) of section 371(c) of this title before the invention thereof by the applicant for patent.

Claims 1, 3-8, 11, 13-15 and 17 are rejected under 35 U.S.C. 102(b) as being clearly anticipated by Faisandier (US 5,800,473).

Faisandier discloses a programmer 300 having a display window on the programmer screen 350.

The examiner is interpreting a first module listing a plurality of cardiac rhythm management system device families to comprise the codes that the programmer 300 compares for implant objects with already known (e.g., an available list of objects in the programmer software 320, which is preferably dynamically managed and safeguarded in a configuration file) (step 30, FIG. 2).

A plurality of second modules is interpreted as comprising the updating of the programmer configuration provided by the connection administrator to complete its software and to adapt the programmer to the implant (steps 50 and 70, FIG. 2).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 2, 9, 10, 12 and 16 are rejected under 35 U.S.C. 103(a) as being unpatentable over Faisandier (US 5,800,473).

Faisandier shows all of the claimed features except for displaying information selected from NGB code.

One of ordinary skill in the art would have found it obvious to include the NGB code for the programmed parametric data because the NGB codes describe the implant parameters needed and because Faisandier teaches most of data of an implant can be contained in a description in the form of a metalanguage, but the recourse to a precompiled code should remain accessible for the integration with original objects.

Response to Arguments

Applicant's arguments filed 5/7/07 have been fully considered but they are not persuasive. Applicant's assertion that Faisandier does not disclose a listing of a plurality of cardiac rhythm management system device families is without merit.


Faisandier teaches the List-Of-objects includes a table containing an object number and an object name. Fig. 1 shows a plurality of families disclosed as a display of software objects in a listing on a user interface display.

Art Unit: 3762

A "module" is a part of a program that performs a distinct function, steps 50 and 70 in Faisandier are distinct functions. Faisandier teaches the display window of programmer 330 has classic display adjustment controls and that the display data 235 or 236 can be displayed in a conventional manner. See col.6, lines 13-17.

Conclusion

Any inquiry concerning this communication or earlier communications from the examiner should be directed to George Manuel whose telephone number is (571) 272-4952.


George Manuel
Primary Examiner
Art Unit: 3762